

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name:	Xpert® Xpress Co	0V-2/Flu/RSV plus		
Cepheid Catalogue P	eart No.: XP3	COV2/FLU/RSV-10		
Kit Lot No.: 1001447	7011			
Cartridge Lot No.: 19	9924			
Kit Expiration Date:	2025-09-07			
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Cepheio Röntge	nvägen 5 54 Solna	Solna Newark	Sunnyvale Lodi IVD (B2)

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

■ If checked this document is produced electronically and valid without a wet signature. Samaneh Vahid Gharavi

Signature of Quality Assurance,		Date
Name:	Samaneh Vahid	
Title	Senior QA Analyst	