

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

	adian Devices Regulations.	curopean
Product Name: Xpert® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10	
Kit Lot No.: 1000364491		
Cartridge Lot No.: 05210		
Kit Expiration Date: 2022	2-12-18	
Legal Manufacturer	Manufacturing Facility OSolna Sunnyvale	
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089	904 Caribbean Drive Sunnyvale, CA 94089	di IVD (B2)
USA	USA	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
If checked this documen	t is produced electronically and valid without a wet signat	ure.
N/h	07/02/2022	
Signature of Quality Assu	rance, Date	
Name: Jenny Nguyen		
Title: Quality Assurance S	upervisor	