



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Norovirus

Catalogue Part No.: GXNOV-CE-10

Kit Lot No.: 1001465061

Cartridge Lot No.: 19501

Kit Expiration Date:2026-04-12

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Functional Testing

| <i>Test Description</i> | <i>Acceptance Criteria</i> | <i>Test Result</i> |
|--------------------------------|--|---------------------------|
| LOW Positive | NORO GI DETECTED;NORO GII DETECTED | Passed |
| HIGH Positive | NORO GI DETECTED;NORO GII DETECTED | Passed |
| Negative | NORO GI NOT DETECTED;NORO GII NOT DETECTED | Passed |

If checked this document is produced electronically and valid without a wet signature.

Anton Engstrom
Anton Engstrom (Dec 27, 2024 10:26 GMT+1)

Signature of Quality Assurance

Date

Name: Anton Andersson Engström

Title: QA Analyst