

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

lian Medical Devices Regulat	ions (CMDR).	
Product Name: Xpert® CT/N	NG	
Cepheid Catalogue Part No.:	GXCT/NG-10	
Kit Lot No.: 1001451818		
Cartridge Lot No.: 42606		
Kit Expiration Date: 2026-1	0-04	
Legal Manufacturer Cepheid 004 Caribbean Drive Sunnyvale, CA 94809 USA		Solna Su
Functional Testing accordi	Acceptance Criteria	Test Result
Positive	CT detected; NG detected	Passed
Negative	CT not detected; NG not detected	Passed
If checked, this document Signature of Quality Assura	is produced electronically and therefore valid withou 10/20/24 unce, Date	it a wet signature
Name: Asia Taylor Torres		