

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpert® X	Apress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part No.:	XP3COV2/FLU/RSV-10		
Kit Lot No.: 1000440329			
Cartridge Lot No.: 03514  Kit Expiration Date: 2023 04	23		
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing	Manufacturing Facility Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	OSolna Newark	Sunnyvale Lodi IVD (B2)

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	

If checked this document is produced electronically and valid without a wet signature.

| Michaelle | 8/29/2012 |
| Signature of Quality Assurance, Date
| Name: Michael Lee

**Quality Systems Specialist** 

Title: