

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Froduct Name: Ape	rtw Apress Cov-2/Flu/RSV plus	
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10	
Kit Lot No.: 1000571801		
Cartridge Lot No.: 21827		
Kit Expiration Date: 2023	3-11-05	
Legal Manufacturer		ınnyvale
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA <i>Functional Testing</i>	Cepheid  904 Caribbean Drive Sunnyvale, CA 94089 USA	odi IVD (B2)
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
Signature of Quality Assu Name: Molly Doan	t is produced electronically and valid without a wet signal and valid without a wet signal arrance, Date	ture.
T:4a. Quality Systems Specialist		