



## CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

**Product Name:** Xpert® Norovirus

**Catalogue Part No.:** GXNOV-CE-10

**Kit Lot No.:** 1001463536

**Cartridge Lot No.:** 19402

**Kit Expiration Date:**2026-04-12

**Legal Manufacturer**

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna,  
Sweden

**Manufacturing Facility**

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna,  
Sweden

***Functional Testing***

<b><i>Test Description</i></b>	<b><i>Acceptance Criteria</i></b>	<b><i>Test Result</i></b>
LOW Positive	NORO GI DETECTED;NORO GII DETECTED	Passed
HIGH Positive	NORO GI DETECTED;NORO GII DETECTED	Passed
Negative	NORO GI NOT DETECTED;NORO GII NOT DETECTED	Passed

If checked this document is produced electronically and valid without a wet signature.

*Samaneh Vahid Gharavi*  
Samaneh Vahid Gharavi (Dec 24, 2024 19:12 GMT+1)

**Signature of Quality Assurance**

**Date**

**Name:** Samaneh Vahid

**Title:** Senior QA Analyst