

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

**Xpert® Xpress CoV-2/Flu/RSV plus Product Name:** 

**Cepheid Catalogue Part No.:** XP3COV2/FLU/RSV-10

Kit Lot No.: 1001455211

Cartridge Lot No.: 30605

**Kit Expiration Date: 2025-10-26** 

Legal Manufacturer **Manufacturing Facility** Solna Sunnyvale Lodi IVD (B2) Newark

Cepheid AB

Röntgenvägen 5 SE-171 54 Solna Sweden

## **Functional Testing**

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

☑ If checked this document is produced electronically and valid without a wet signature. Alexander Avramidis

Signature of Quality Assurance,	Date
Name: Alexander Avramidis	
Title: QA Analyst	