



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Bladder Cancer Monitor

Cepheid Catalogue Part No.: GXBLAD-CM-CE-10

Kit Lot No.: 1001465222

Cartridge Lot No.: 22103

Kit Expiration Date: 2026 12 27

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

Solna

Sunnyvale

Lodi

Functional Testing according to D22403, Rev. G

| <i>Test Description</i> | <i>Acceptance Criteria</i> | <i>Test Result</i> |
|-------------------------|----------------------------|--------------------|
| Negative | Negative | Passed |
| LOW Positive | Positive | Passed |
| HIGH Positive | Positive | Passed |

If checked, this document is produced electronically and therefore valid without a wet signature



Signature of Quality Assurance

13 Jan 25

Date

Name: Chong Moua

Title: Quality Assurance Specialist