

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.

Test Description	Acceptance Criteria			Test Result		
Functional Testing	according to D3699	0, Rev. E.1				
		USA	\odot	Lodi		
USA		Lodi, CA 95240				
Drive Sunnyvale, C.	A 94089	121 N Guild Avenue		Some		,
Cepheid 904 Caribb	ean	Cepheid		Solna		Sunnyvale
Legal Manufacture	er	Manufacturing Facility				
Kit Expiration Date:	2025 10 12					
Cartridge Lot No.:	03103					
Kit Lot No.: 100146	54475					
Cepheid Catalogue F	Part No.: XPRSMVP	-10				
Product Name: Xp	ert® Xpress MVP					

Test Description	Acceptance Criteria	Test Result	
Positive	BV POSITIVE; Candida group DETECTED; Candida glab-krus DETECTED; TV DETECTED	Passed	
Negative	BV NEGATIVE; Candida group NOT DETECTED; Candida glab-krus NOT DETECTED; TV NOT DETECTED	Passed	

If checked, this document is produced electronically and therefore valid without a wet signature

Signature of Quality Assurance

Date

12/20/2024

Name: Mark Magno Shelor

Title: Quality Assurance Specialist