CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert Xpress SARS-CoV-2

Cepheid Catalogue Part No.: XPRSARS-COV2-10

Kit Lot No.: 1001467018

Cartridge Lot No.: 39204

Kit Expiration Date: 2026-06-28

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility (•) Solna ( ) Sunnyvale ( ) Newark

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Functional Testing

Test Description	*Acceptance Criteria*	*Test Result*
Negative	SARS-CoV-2 NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE	Passed

☐ If checked this document is produced electronically and valid without a wet signature.

Anton Andersson 2025-01-10

Signature of Quality Assurance **Date**

Name: Anton Andersson Engström

Title: QA Analyst