

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: X <sub>I</sub>	pert® Carb	a-R				
Cepheid Catalogue	Part No.:	GXCARBAR-CN-10	)			
Kit Lot No.: 10014	36607					
Cartridge Lot No.:	23002					
<b>Kit Expiration Date</b>	: 2026-01	-04				
Legal Manufactur Cepheid 904 Caribbean Driv Sunnyvale, CA 940  Functional Testing	 89 USA	904 Ca	ribbean Drive vale, CA 94089		Solna	Sunnyvale
Test Description		Acceptance Criteria			Test Result	$\neg$
Low Positive	IMP DETECT	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED			Passed	_
High Positive	IMP DETECT	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED			Passed	
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED			Passed		
If checked, this of Molly Doan  Molly Doan  Molly Doan (Jul 23, 2024 09:34 PDT)	locument is	s produced electronical	ly and therefore valid Jul 23, 2024	d without a	a wet signature	
Signature of Quali	ty Assuran	ce,	Date	_		

Name: Molly Doan

Title: Quality Systems Specialist

## 301-6258 Rev C CofA GXCARBAR

Final Audit Report 2024-07-23

Created: 2024-07-23

By: Molly Doan (molly.doan@cepheid.com)

Status: Signed

Transaction ID: CBJCHBCAABAAmbAmEGwqK06h3qaUPQGt3Xi1kgb05l7p

## "301-6258 Rev C CofA GXCARBAR" History

Document created by Molly Doan (molly.doan@cepheid.com) 2024-07-23 - 4:32:24 PM GMT- IP address: 45.62.187.253

Document emailed to Molly Doan (molly.doan@cepheid.com) for signature 2024-07-23 - 4:34:24 PM GMT

Document e-signed by Molly Doan (molly.doan@cepheid.com)

Signature Date: 2024-07-23 - 4:34:50 PM GMT - Time Source: server- IP address: 45.62.187.253

Agreement completed. 2024-07-23 - 4:34:50 PM GMT

