



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Carba-R

Cepheid Catalogue Part No.: GXCARBAR-CN-10

Kit Lot No.: 1001160049

Cartridge Lot No.: 22004

Kit Expiration Date: 2025-02-09

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA

Solna

Sunnyvale

Lodi

Functional Testing according to D18272, Rev. AE

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Low Positive	IMP DETECTED; VIM DETECTED; NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
High Positive	IMP DETECTED; VIM DETECTED; NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

Pauline Vu 08/29/2023
Signature of Quality Assurance, Date

Name: Pauline Vu

Title: Quality Systems Specialist