

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpe	ert® Xpress CoV-2/Flu/RSV plus	
Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10		
Kit Lot No.: 1000604922		
Cartridge Lot No.: 14714		
Kit Expiration Date: 2023-11-26		
Legal Manufacturer		nnyvale
Cepheid	Cepheid AB Newark OLo	di IVD (B2)
904 Caribbean Drive	Röntgenvägen 5	
Sunnyvale, CA 94089	SE-171 54 Solna	
USA	Sweden	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
☐ If checked this documen	t is produced electronically and valid without a wet signature	ure.
Saraph	2022-12-07	
Signature of Quality Assurance, Date		
Name: Sara Mustafa Abdul	la	
Title: QA Analyst		