

## CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Prod	uct	Nam	e:
	acc	T I COMM	

**Xpert® Xpress CoV-2/Flu/RSV plus** 

Cepheid Catalogue Part No.:

XP3COV2/FLU/RSV-10

Kit Lot No.: 1001455040

Cartridge Lot No.: 63218

Kit Expiration Date: 2025-10-26

Legal Manufacturer

**Manufacturing Facility** 

Solna Newark

Sunnyvale Lodi IVD (B2)

Cepheid

904 Caribbean Drive

Sunnyvale, CA 94089

**USA** 

Cepheid

121 N Guild Avenue

Lodi, CA 95240

**USA** 

Functional Testing

Test Description	tion Acceptance Criteria	
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature. 07 NOV 2024

Date

Signature of Quality Assurance

Name: Karla Camarena

Title: Quality Systems Specialist

PN 302-7224 Rev. C