

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

IVD Directive and the Can	nadian Devices Regulations.	
Product Name: Xpert® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10	
Kit Lot No.: 100145896	6	
Cartridge Lot No.: 3071	0	
Kit Expiration Date: 202	25-11-16	
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing		nnyvale di IVD (B2)
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
■ If checked this document Lava Maroof Lava Maroof Lava Maroof (Nov 27, 2024 11:47 GMT+1) Signature of Quality Assort Name: Lava Maroof	nt is produced electronically and valid without a wet signaturance, Date	ire.
Title: QA Analyst		