

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA					
Cepheid Catalogue Part No.: GXMRSA-	100N-10 🔽				
Kit Lot No.: 1000878025					
Cartridge Lot No.: 23825					
Kit Expiration Date: 2025-02-23					
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA	0	olna	0	Sunnyvalo
Functional Testing according to D1689	o, Kev. AA				

Test Description	Acceptance Criteria	Test Result	
Positive	MRSA Positive	Passed	
Negative	MRSA Negative	Passed	

If checked, this document is produced electronically and therefore valid without a wet signature

Signature of Quality Assurance,

Name: Julie Millet

Title: Quality Systems Specialist