

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR) and the China National Medical Products Administration's Quality System Requirements.

**Product Name:** Xpert® MTB/RIF

Catalogue Part No.: GXMTB/RIF-CN-50

Kit Lot No.: 1001467998

Cartridge Lot No.: 95106

Kit Expiration Date: 2027-01-10

Legal Manufacturer Manufacturing Facility

Cepheid ABCepheid ABRöntgenvägen 5Röntgenvägen 5SE-171 54 Solna,SE-171 54 Solna,

Sweden Sweden

Functional Testing

Test Description	Acceptance Criteria	Test Result
Positive	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED or MTB DETECTED LOW; Rif Resistance NOT DETECTED or MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED or MTB DETECTED HIGH; Rif Resistance NOT DETECTED	Passed
Negative	MTB NOT DETECTED	Passed

✓ If checked this document is produced electronically and valid without a wet signature.

Sara Mustafa Abdulla Sara Mustafa Abdulla (Jan 28, 2025 12:39 GMT+1)	

Name: Sara Mustafa Abdulla

Title: QA Analyst