



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid’s Quality System, in compliance with the US Food and Drug Administration’s Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Check

Cepheid Catalogue Part No.: XPERTCHECK-CE-5

Kit Lot No.: 1001456667

Cartridge Lot No.: 06512

Kit Expiration Date: 2025-09-08

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA



Solna



Sunnyvale



Lodi

Functional Testing according to D14336 Rev. R

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Functional Test	Testing results meet specifications	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

12/19/24

Signature of Quality Assurance,

Date

Name: Mai Kue

Title: Quality System Specialists