



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.

Product Name: Xpert® Mpx

Cepheid Catalogue Part No.: GXMPX-10

Kit Lot No.: 1001447295

Cartridge Lot No.: 01201

Kit Expiration Date: 2025-09-14

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA

Solna


Sunnyvale

Lodi

Functional Testing according to D61298, Rev. D

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MPXV Clade II Detected	Passed
Negative	MPXV Clade II Not Detected; OPXV Not Detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature


Signature of Quality Assurance

26 Sep 24
Date

Name: Molly Doan

Title: Quality Systems Specialist