

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the Japanese Pharmaceuticals and Medical Devices Law, ISO 13485 and the European IVD Directive.

Product	Name:	3
1 I Uuuct	Maille.	1

Xpert® MTB/RIF

Cepheid Catalogue Part No.:

GXMTB/RIF-JP-10

Kit Lot No.:

1001458870

Cartridge Lot No.:

94601

Kit Expiration Date:

2026-08-02

Legal Manufacturer

Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden **Manufacturing Facility**

Cepheid AB Röntgenvägen 5 SE-17154 Solna

Sweden

Functional Testing according to D66612, Rev. A

Test Description	Acceptance Criteria	Test Result
Mutant Strain 1-High concentration ((MS1(H))	MTB DETECTED; Rif Resistance DETECTED	Passed
Mutant Strain 1-Low concentration ((MS1(L))	MTB DETECTED; Rif Resistance DETECTED	Passed
Mutant Strain 2-High concentration ((MS2(H))	MTB DETECTED; Rif Resistance DETECTED	Passed
Mutant Strain 2-Low concentration ((MS2(L))	MTB DETECTED; Rif Resistance DETECTED	Passed
Wild Type (WT) TB	MTB DETECTED; Rif Resistance NOT DETECTED	Passed
Negative Control	MTB NOT DETECTED	Passed

Final Judgement

2 111111 0 11118 0 111	1 titut o titigement		
Product Specification	Judgement		
Sensitivity	Passed		
Precision	Passed		
Simultaneous Repeatability	Passed		

☐ If checked this document is produced electronically and valid without a wet signature.

LOW	ORT FIEDUSE	20241129		
Signature of Quality Assurance, Date				
Name:	Robert Fiedler			

Title:_QA Analyst