

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

- Sanda Carlo Carl	adian Devices Regulations.	Buropean	
Product Name: Xpe	oduct Name: Xpert® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10	**	
Kit Lot No.: 1000541856 Cartridge Lot No.: 21805 Kit Expiration Date: 2023		¥.	
•			
Legal Manufacturer			
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Cepheid Newark Lo 904 Caribbean Drive Sunnyvale, CA 94089 USA	di IVD (B2)	
Functional Testing			
Test Description	Acceptance Criteria	Test Result	
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed	
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed	
If checked this documents	tis produced electronically and valid without a wet signat	ure.	
Name: Molly Doan			
Title: Quality Systems Spe	ecialist		