



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® C.difficile

Cepheid Catalogue Part No.: GXCDIFFICILE-10

Kit Lot No.: 1001459079

Cartridge Lot No.: 41601

Kit Expiration Date: 2026-05-17

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

Solna

Sunnyvale

Lodi

Functional Testing according to D37468, Rev. E.1

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	Toxigenic C.diff POSITIVE	Passed
Negative	Toxigenic C.diff NEGATIVE	Passed

If checked, this document is produced electronically and therefore valid without a wet signature



Signature of Quality Assurance,

12/01/24

Date

Name: Kimberly Perez

Title: Quality Systems Specialist II