

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® Carba-R
Cepheid Catalogue Part No.: GXCARBAR-CE-10
Kit Lot No.: 1000602642
Cartridge Lot No.: 20707
Kit Expiration Date: 2024-05-19

Legal Manufacturer
Cepheid

904 Caribbean Drive Sunnyvale, CA 94089 USA Manufacturing Facility

Solna

Sunnyvale

Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP1 NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

Uhren	12-03-22	
Signature of Quality Assurance	Date	
Name: Renee Argishti		
Title: Sr. Quality Systems Specialist		