

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xper	t® MTB/RIF Ultra	
Cepheid Catalogue P	art No.: GXMTB/RIF-ULTRA-50	
Kit Lot No.: 100060		
Cartridge Lot No.: 4	4212	
Kit Expiration Date:	2023 11 05	
Legal Manufacture Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA) Solna Sunnyva
	according to D25862, Rev. A E	Test Result
Test Description	Acceptance Criteria	1est Kesuit
Wild Type Control	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED OT MTB DETECTED LOW; Rif Resistance NOT DETECTED OF MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED OF MTB DETECTED HIGH; Rif Resistance NOT DETECTED	Passed
Mutant Control	MTB DETECTED VERY LOW; RIF Resistance DETECTED or MTB DETECTED LOW; RIF Resistance DETECTED or MTB DETECTED MEDIUM; RIF Resistance DETECTED or MTB DETECTED HIGH; RIF Resistance DETECTED	Passed
Negative	MTB NOT DETECTED	Passed
If checked, this do	cument is produced electronically and therefore valid without 2 20 2022 Assurance, Date	a wet signature
Name: Mai Kue		