CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA

Cepheid Catalogue Part No.: GXMRSA-100N-10

Kit Lot No.: 1001431183

Cartridge Lot No.: 25201

Kit Expiration Date: 2026 05 17

Legal Manufacturer
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility
Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

○ Solna ○ Sunnyvale

◉ Lodi

Functional Testing according to D16895, Rev. AA

Test Description	*Acceptance Criteria*	*Test Result*
Positive	MRSA Positive	Passed
Negative	MRSA Negative	Passed

☐ If checked, this document is produced electronically and therefore valid without a wet signature

02 Jun 2024

Signature of Quality Assurance, **Date**

Name: Nancy Inthavisak

Title: Quality Systems Specialist