



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Carba-R

Catalogue Part No.: Choose One

Kit Lot No.: 1001455555

Cartridge Lot No.: 12202

Kit Expiration Date: 2026-04-19

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Solna

Sunnyvale

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

If checked this document is produced electronically and valid without a wet signature.

Samaneh Vahid Gharavi
Samaneh Vahid Gharavi (Oct 31, 2024 15:29 GMT+1)

Signature of Quality Assurance

Date

Name: Samaneh Vahid

Title: Senior QA Analyst