

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

<b>Product Name:</b>	Xpert(
Froduct Name:	Aperto

® Xpress CoV-2 plus

Cepheid Catalogue Part No.:

XP3SARS-COV2-10

Kit Lot No.:

1001455064

Cartridge Lot No.:

21009

**Kit Expiration Date:** 

2025 10 26

Legal	Ma	nufa	ctur	er
LICEA.	1 1110	mula	ctui	<u>UI</u>

Manufacturing Facility Solna Sunnyvale Newark Lodi Cepheid

Cepheid

904 Caribbean Drive

121 N Guild Ave

Sunnyvale, CA 94089 USA

Lodi, CA 95240 USA

## **Functional Testing**

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance,

Name: Trinh Nguyen

Title: Quality Systems Specialist