



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA

Cepheid Catalogue Part No.: GXMRSA-100N-10

Kit Lot No.: 1001426479

Cartridge Lot No.: 25002

Kit Expiration Date: 2026 04 19

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

Solna


Sunnyvale

Lodi

Functional Testing according to D16895, Rev. AA

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MRSA Positive	Passed
Negative	MRSA Negative	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

 05/10/2024

Signature of Quality Assurance, Date

Name: Crystal Sysenglath

Title: Quality Systems Specialist