

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.

Product Name: Xpert® Xpr Catalogue Part No.: XPRS4 Kit Lot No.: 1001448200 Cartridge Lot No.: 62003 Kit Expiration Date: 2025 0	Part No. and Rev	
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing according	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA Sta D47377 Pay I	Sunnyvale
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
☐ If checked this document	is produced electronically and valid without a wet signature. $09/25/2024$	
Signature of Quality Assurance, Date		
Name: Asia Taylor Torres Title: Quality Systems Spec	rialist	