



## CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

**Product Name:** Xpert® SA Nasal Complete

**Catalogue Part No.:** GXSACOMP-CE-10

**Kit Lot No.:** 1001451724

**Cartridge Lot No.:** 16402

**Kit Expiration Date:** 2026-07-26

**Legal Manufacturer**

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna  
Sweden

**Manufacturing Facility**

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna  
Sweden



Solna



Sunnyvale

***Functional Testing***

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MRSA POSITIVE; SA POSITIVE	Passed
Negative	MRSA NEGATIVE; SA NEGATIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

2024-10-18

Signature of Quality Assurance

Date

**Name:** Alexander Avramidis

**Title:** QA Analyst

PN 301-6241 Rev. D