

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the cartridges specified below conform to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MTB/RIF Cartridges

Cepheid Part No.: 700-7773 - ASSY,10 CART POUCH,MTB/RIF BANGALORE

Lot No.: 1001470356

Cartridge Lot No.: 95119

Manufacuring Date: 2025-01-31

**Expiration Date:** 2027-01-17

Manufacturing Facility Solna Sunnyvale Newark

Cepheid AB Röntgenvägen 5 SE-171 54 Solna

Sweden

## **Functional Testing**

Test Description	Acceptance Criteria	Test Result
Positive	MTB DETECTED VERY LOW;Rif Resistance NOT DETECTED	1 a
	or MTB DETECTED LOW;Rif Resistance NOT DETECTED	-
	or	Passed
	MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED	- 1
4 2	or	10
	MTB DETECTED HIGH; Rif Resistance NOT DETECTED	
Negative	MTB NOT DETECTED	Passed

L 2025-02-07 Signature of Quality Assurance

Date

Name: Alexander Avramidis

Title: QA Analyst