

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® M1B/RIF Ultra	ı		
Cepheid Catalogue Part No.: GXMTB/	RIF-ULTRA-50		
Kit Lot No.: 1001465206			
Cartridge Lot No.: 61402			
Kit Expiration Date: 2026-03-22			
Legal Manufacturer	Manufacturing Facility		
Cepheid AB	Cepheid	Solna	Sunnyvale
Röntgenvägen 5	121 N Guild Avenue	o Bonna	Sumiy vale
SE-17154 Solna	Lodi, CA 95240		
Sweden	USA	Lodi	

Functional Testing according to D25862, Rev. AP

Test Description	Description Acceptance Criteria	
Wild Type Control	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED or MTB DETECTED LOW; Rif Resistance NOT DETECTED or MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED or MTB DETECTED HIGH; Rif Resistance NOT DETECTED	Passed
Mutant Control	MTB DETECTED VERY LOW;RIF Resistance DETECTED or OF MTB DETECTED LOW;RIF Resistance DETECTED OF MTB DETECTED DETECTED OF MTB DETECTED MEDIUM;RIF Resistance DETECTED OF MTB DETECTED HIGH;RIF Resistance DETECTED	Passed
Negative	Negative MTB NOT DETECTED	

Mutan	WILLIAM COMMON MTB DETECTED MEDIUM; RIF Resistance DETECTED or MTB DETECTED HIGH; RIF Resistance DETECTED					
N	legative	e MTB NOT DETECTED				
If checked, this document is produced electronically and therefore valid without a wet signature						
J	Mai Kue	y Assurance,	Date			
Title:	Ouality Sys	tem Specialist				