



**CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid’s Quality System, in compliance with the US Food and Drug Administration’s Quality System Requirements and ISO 13485.

**Product Name:** Xpert® Xpress CoV-2/Flu/RSV *plus*

**Catalogue Part No.:** XPRS4PLEX-10

**Kit Lot No.:** 1001462618

**Cartridge Lot No.:** 64405

**Kit Expiration Date:** 2025-12-07

**Instructions for Use (IFU)**

**Part No. and Rev.:**

302-8057  Rev  A

302-8055 Rev  C

303-1854 Rev  B

**Legal Manufacturer**

Cepheid  
904 Caribbean Drive  
Sunnyvale, CA  
94089 USA

**Manufacturing Facility**

Cepheid  
904 Caribbean Drive  
Sunnyvale, CA 94089  
USA

Solna  
 Lodi

Sunnyvale

***Functional Testing according to D47377, Rev. L***

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

\_\_\_\_\_  
Signature of Quality Assurance,

18 Dec 24  
\_\_\_\_\_  
Date

**Name:** Molly Doan

**Title:** Quality Systems Specialist