



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations.

Product Name: Xpert® BCR-ABL Ultra p190

Cepheid Catalogue Part No.: GXBCRABLP190-CE-10

Kit Lot No.: 1000486342

Cartridge Lot No.: 01801

Kit Expiration Date: 2023-10-29

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA

Solna Sunnyvale Newark Lodi

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Negative	NEGATIVE [Sufficient ABL transcript]	Passed
~0.02% (p190/ABL)	POSITIVE	Passed
~0.1% (p190/ABL)	POSITIVE	Passed


Signature of Quality Assurance,

08/29/2022
Date

Name:

Title: