

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Pro	dn	cŧ	Na	m	۵٠
1 1 1 1	uu	CL.	1 2		С.

Xpert® Xpress CoV-2/Flu/RSV plus

Cepheid Catalogue Part No.:

XP3COV2/FLU/RSV-10

Kit Lot No.: 1000415657

Cartridge Lot No.: 07404

Kit Expiration Date: 2023-05-21

Legal Manufacturer

Manufacturing Facility

Sunnyvale
Lodi IVD (B2)

)Solna)Newark

Cepheid

Cepheid

904 Caribbean Drive

904 Caribbean Drive Sunnyvale, CA 94089

Sunnyvale, CA 94089 **USA**

USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance, Date

Name: Emaela S. Quan-Tong

Title: Quality Assurance Supervisor