

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

1 1 Directive and the Canadian Devices Regulations.			
Product Name: Xpert® Xpress CoV-2/Flu/RSV plus			
Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10			
Kit Lot No.: 1000559125			
Cartridge Lot No.: 09801			
Kit Expiration Date: 2023 10 29			
Legal Manufacturer	Manufacturing Facility		unnyvale
Cepheid	Cepheid	Newark OL	odi IVD (B2)
904 Caribbean Drive	904 Caribbean Drive		
Sunnyvale, CA 94089	Sunnyvale, CA 94089		
	USA		
USA	OSA		
Functional Testing			
Test Description	Acceptance Criteria		Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B		Passed
NEGATIVE;RSV NEGATIVE			
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE		Passed
If checked this document is produced electronically and valid without a wet signature.			
paulnerla			
Signature of Quality Assurance, Date			
Name: Pauline Vu			
Title: Quality Systems Specialist			