

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations.

Product Name:

Xpert® BCR-ABL Ultra p190

Cepheid Catalogue Part No.:

GXBCRABLP190-CE-10

Kit Lot No .:

1001416944

Cartridge Lot No.:

02601

Kit Expiration Date:

2025-04-06

Legal Manufacturer

Manufacturing Facility Solna Sunnyvale Newark Lodi

Cepheid

Cepheid

904 Caribbean Drive

904 Caribbean Drive

Sunnyvale, CA 94089 USA

Sunnyvale, CA 94089

USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	NEGATIVE [Sufficient ABL transcript]	Passed
~0.02% (p190/ABL)	POSITIVE	Passed
~0.1% (p190/ABL)	POSITIVE	Passed

Signature of Quality Assurance.

Name: Roya Kazemi

Title:

Quality Systems Specialist