Cepheid®

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® HIV-1 Viral Load XC

Cepheid Catalogue Part No.: GXHIV-VL-XC-CE-10

Kit Lot No.: 1001426361

Cartridge Lot No.: 15508

Kit Expiration Date: 2025-10-12

Legal Manufacturer
Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility
Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

(●) Solna ( ) Sunnyvale
( ) Lodi

Functional Testing according to D51111, Rev. C

Test Description	*Acceptance Criteria*	*Test Result*
HIV-1 VL XC PC Low	HIV-1 DETECTED: 2.08-2.52 log10 IU/mL	Passed
HIV-1 VL XC PC High	HIV-1 DETECTED: 5.54-5.86 log10 IU/mL	Passed
Negative	HIV-1 NOT DETECTED	Passed

☐ If checked, this document is produced electronically and therefore valid without a wet signature

Ron 20240430

Signature of Quality Assurance **Date**

Name: Robert Fiedler

Title: QA Analyst