

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_	Xpert MTB/XDR	
Cepheid Catalogue I	Part No.: GXMTB/XDR-10	
Kit Lot No.: 100	00874073	
Cartridge Lot No.:0)2403	
Kit Expiration Date	:2025-03-09	
Legal Manufacturer		Manufac

Manufacturing Facility Solna
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Sunnyvale

Functional Testing

Cepheid AB

Sweden

Röntgenvägen 5

SE-17154 Solna

Test Description	` Acceptance Criteria	Test Result
Negative	MTB NOT DETECTED	Passed
Positive - Wild Type	MTB DETECTED;INH Resistance NOT DETECTED;FLQ Resistance NOT DETECTED;AMK Resistance NOT DETECTED;KAN Resistance NOT DETECTED;CAP Resistance NOT DETECTED;ETH Resistance NOT DETECTED	Passed
Positive - Mutant	MTB DETECTED;INH Resistance DETECTED;FLQ Resistance DETECTED;AMK Resistance DETECTED;KAN Resistance DETECTED;CAP Resistance DETECTED;ETH Resistance DETECTED	Passed

	1	22120122,21111000
Mulh	Mc	24Apr.23
Signature of Quality	ssurance	Date
[
Name: Molly Doan	1	
Title: Quality Syste	ems Specialist	