

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

P	ro	du	ct	N	am	e:	
	LU	uu	·	TI	am	· ·	

**Xpert® Xpress Flu/RSV** 

Cepheid Catalogue Part No.: XPRSFLU/RSV-10



Solna

Kit Lot No.: 1001469003

Cartridge Lot No.: 46514

Kit Expiration Date: 2026-10-18

Legal Manufacturer

**Manufacturing Facility** 

Sunnyvale

Cepheid

Cepheid

121 N Guild Avenue

904 Caribbean Drive Sunnyvale, CA 94089

Lodi, CA 95240

USA

**USA** 

## **Functional Testing**

Test Description	Acceptance Criteria	Test Result Passed
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance,

Date

Name: Karla Camarena

Title: Quality Systems Specialist