

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name:	Xpert® Xpress	CoV-2/Flu/RSV plus	5

Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10

Kit Lot No.: 1001465264

Cartridge Lot No.: 64219

Kit Expiration Date: 2026-02-01

Legal ManufacturerManufacturing FacilityOsolnaSunnyvaleCepheidNewarkLodi IVD (B2)

904 Caribbean Drive 121 N Guild Avenue

Sunnyvale, CA 94089 Lodi, CA 95240

USA USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

and than 09FEB2025

Signature of Quality Assurance, Date

Name: Aaron Shaw

Title: Quality Systems Specialist I