

Product Name:

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Xpert® Xpress CoV-2/Flu/RSV plus

^	•		
Cepheid Catalogue Part No	o.: XP3COV2/FLU/RSV-	-10	
Kit Lot No.: 1001357662			
Cartridge Lot No.: 42602			
Kit Expiration Date: 2024-	11-17		
Legal Manufacturer	Manufacturing Facility	OSolna	Sunnyvale
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Cepheid 121 N Guild Avenue Lodi, CA 95240 USA	Newark	Lodi IVD (B2)
Functional Testina			

Test Description Acceptance Criteria		Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

29 NOV 2023

Signature of Quality Assurance,

Date

Name: Sayed Safdari

Title: Quality Systems Specialist I