

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

| Product Name: Xpe | rt® Xpress CoV-2/Flu/RSV plus | | | | | |
|--|---|------------------------|--|--|--|--|
| Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10 | | | | | | |
| Kit Lot No.: 1000366950 Cartridge Lot No.: 05223 Kit Expiration Date: 2022-12-25 | | | | | | |
| Legal Manufacturer | Newark OLO | nnyvale di IVD (B2) | | | | |
| Cepheid | Cepheid | di IVD (B2) | | | | |
| 904 Caribbean Drive | 904 Caribbean Drive | | | | | |
| Sunnyvale, CA 94089 | Sunnyvale, CA 94089 | | | | | |
| USA | USA | | | | | |
| Functional Testing | • | | | | | |
| Test Description | Acceptance Criteria | Test Result | | | | |
| Negative | SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE | Passed | | | | |
| Positive | SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | Passed | | | | |
| ☐ If checked this document is produced electronically and valid without a wet signature. | | | | | | |

| X | V/1- | 67 | 02/20 | 22 |
|--------|------------------------------|----|-------|----|
| Signat | ure of Quality Assurance, | | Date | |
| Name: | Jenny Nguyen | | | |
| 7F:41 | Quality Assurance Supervisor | | | |