

Signature of Quality Assurance,

Title: Quality System Specialist

Name: Austin Mei

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

1 v D Directive and the Canadian Devices Regulations.			
Product Name: Xpert® Xpress CoV-2/Flu/RSV plus			
Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10			
Kit Lot No.: 1000688834			
Cartridge Lot No.: 26212			
Cartridge Lot 110 20212			
Kit Expiration Date: 2024-01-07			
Legal Manufacturer	Manufacturing Facility		ınnyvale
Cepheid	Cepheid	Newark OLo	odi IVD (B2)
904 Caribbean Drive	904 Caribbean Drive		
Sunnyvale, CA 94089	Sunnyvale, CA 94089		
USA	USA		
Functional Testing			
Test Description	Acceptance Criteria		Test Result
Negative	SARS-CoV-2 NEGATIVE; Flu A NEGATIVE; Flu B		Passed
	NEGATIVE;RSV NEGATIVE		
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE		Passed
☐ If checked this document is produced electronically and valid without a wet signature.			