

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_ Xpert MTB/XDR			
Cepheid Catalogue Part No.: GXMTB/XDR-10			
Kit Lot No.: 1001413438			
Cartridge Lot No.: 04102			
Kit Expiration Date: 2026-01-11			
Legal Manufacturer	Manufacturing Facility	○ Solna	Sunnyvale
Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden	Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA		

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	MTB NOT DETECTED	Passed
Positive - Wild Type	MTB DETECTED;INH Resistance NOT DETECTED;FLQ Resistance NOT DETECTED;AMK Resistance NOT DETECTED;KAN Resistance NOT DETECTED;CAP Resistance NOT DETECTED;ETH Resistance NOT DETECTED	Passed
MTB DETECTED;INH Resistance DETECTED;FLQ Resistance DETECTED;AMK Resistance DETECTED;KAN Resistance DETECTED;CAP Resistance DETECTED;ETH Resistance DETECTED		Passed

Theyhur	02/06/2024
Signature of Quality Assurance	Date
V	
Name: ThuyTien Nguyen	
Title: Quality Systems Specialist	