

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xper	t® MTB/RIF Ultra		
Cepheid Catalogue Part No.: GXMTB/RIF-ULTRA-50			
Kit Lot No.: 100141	7594		
Cartridge Lot No.: 4	7309		
Kit Expiration Date:	2025 08 24		
Legal Manufacture			
Cepheid AB	Cepheid) Solna (Sunnyva	
Röntgenvägen 5	121 N Guild Avenue		
SE-17154 Solna	Lodi, CA 95240) Lodi	
Sweden	USA) Loui	
Functional Testing	according to D25862, Rev. AN		
Test Description	Acceptance Criteria	Test Result	
Wild Type Control	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED of MTB DETECTED LOW; Rif Resistance NOT DETECTED of MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED of MTB DETECTED HIGH; Rif Resistance NOT DETECTED	Passed	
Mutant Control	MTB DETECTED VERY LOW;RIF Resistance DETECTED or MTB DETECTED LOW;RIF Resistance DETECTED or MTB DETECTED MEDIUM;RIF Resistance DETECTED or MTB DETECTED HIGH;RIF Resistance DETECTED	Passed	
Negative	MTB NOT DETECTED	Passed	
If checked, this do	cument is produced electronically and therefore valid without	a wet signature	
Sand Brogge sol	10Mar 2024	5	
	Negative MTB NOT DETECTED Negative MTB NOT DETECTED Passed If checked, this document is produced electronically and therefore valid without a wet signature Signature of Quality Assurance, Date		
Name: Sarah Bagas	col		
Title: Quality Ass	urance Specialist		