

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

<b>Product Name:</b>	Xpert® Xpress	s CoV-2/Flu/RSV plus
----------------------	---------------	----------------------

Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10

Kit Lot No.: 1001456669

Cartridge Lot No.: 63322

Kit Expiration Date: 2025 11 02

Legal ManufacturerManufacturing FacilityOsolnaSunnyvaleCepheidCepheidNewarkLodi IVD (B2)

Cepheid Cepheid 904 Caribbean Drive 121 N Guild Avenue

Sunnyvale, CA 94089 Lodi, CA 95240

USA USA

**Functional Testing** 

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance, Date

Name: Jennifer Nguyen

Title: Quality Systems Specialist