

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Cepheid Catalogue	Part No ·	GXCARBA	R-CN-10					
Kit Lot No.: 1001	370425							
Cartridge Lot No.:	21216							
Kit Expiration Date	2025-05-	11						
Legal Manufactur	<u>er</u>		<u>Manufactur</u>	ing Facility				
Cepheid			Cepheid	ъ.		Solna	•	Sunnyvale
904 Caribbean Drive Sunnyvale, CA 94089 USA			904 Caribbean Drive Sunnyvale, CA 94089		_		_	
unity vale, CA 940	109 USA		USA	A 94009	0	Lodi		
Functional Testing	according	g to D18272	, Rev. AF			ā		
Test Description		Acceptance Criteria				Test Result		7
Low Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED			ETECTED	Passed			
High Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED			ETECTED	Passed],	
Negative	IMP NOT DETECTED, VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED, OXA4\$ NOT DETECTED			DETECTED	Passed]′	

Name: Zenny Santos

Title: Quality Systems Specialist