

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR) and the China National Medical Products Administration's Quality System Requirements.

System Requirements.	is (CMDR) and the China National Medical Products Adi	ninistration's Qualit
Product Name: Xpert® CT/N	1G	
Catalogue Part No.: GXCT	VNGX-CE-10   ▼	
Kit Lot No.: 1001449530		
Cartridge Lot No.: 36109		
Kit Expiration Date: 2026-0	)9-20	
Legal Manufacturer Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden  Functional Testing	Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	Sunnyvale
Test Description	Acceptance Criteria	Test Result
Positive	CT DETECTED; NG DETECTED	Passed
Negative	CT NOT DETECTED; NG NOT DETECTED	Passed
If checked this document is	is produced electronically and valid without a wet signature.	
Signature of Quality Assurance Date		

Name: Lava Maroof

Title: QA Analyst

PN 301-6236 Rev. E